SENSIDYNE®
Precision Gas Detector Tubes

OPEN

Tube No. 106SC
CARBON MONOXIDE
1–50ppm
10 Tubes
Lot No.862022

Store in a cool and dark place

Made in Japan nSL10040

For use with Sensidyne piston-type gas detection pumps only.

CAUTION:
- Leak test pump before use.
- Read enclosed instructions before use.

For technical support call:
Tel: 800.451.9444 / 727.530.3602
Fax: 727.539.0550
E-mail: Info@Sensidyne.com

Sensidyne, LP • 1000 112th Circle North, Suite 100
St. Petersburg, FL 33716 USA • www.Sensidyne.com

SENSIDYNE Precision Gas Detector Tubes



KITAGAWA
KOMYO RIKAGAKU KOGYO

INSTRUCTION MANUAL
CARBON MONOXIDE DETECTOR TUBE
No.106SC

★ READ CAREFULLY THIS INSTRUCTION MANUAL AND THE INSTRUCTIONS OF THE ASPIRATING PUMP PRIOR TO USING THIS PRODUCT.
★ DO NOT DISCARD THIS INSTRUCTION MANUAL UNTIL ALL THE TUBES IN THIS BOX ARE USED UP.

1. PERFORMANCE:

Measuring Range	: 1 - 50 ppm
and Pump Stroke	: 1 pump stroke
Sampling Time	: 4 minutes

SENSIDYNE®
Precision Gas Detector Tubes

OPEN

Tube No. 103SE

SULPHUR DIOXIDE

0.25–10ppm

10 Tubes

Lot No.233112

Store in refrigerated place at 32~50°F(0~10°C)

Made in Japan nSL10350

For use with Sensidyne piston-type gas detection pumps only.

CAUTION:

- Leak test pump before use.
- Read enclosed instructions before use.

For technical support call:
Tel: 800.451.9444 / 727.530.3602
Fax: 727.539.0550
E-mail: info@Sensidyne.com

Sensidyne, LP • 1000 112th Circle North, Suite 100
St. Petersburg, FL 33716 USA • www.Sensidyne.com

Precision Gas Detector Tubes

100mL

Tube No.103SE SO_2

100mL

Tube No.103SE SO_2

100mL

KITAGAWA
KOMYO RIKAGAKU KOGYO

INSTRUCTION MANUAL
SULPHUR DIOXIDE DETECTOR TUBE

No.103SE

★ READ CAREFULLY THIS INSTRUCTION MANUAL AND THE INSTRUCTIONS OF THE ASPIRATING PUMP PRIOR TO USING THIS PRODUCT.
★ DON'T DISCARD THIS INSTRUCTION MANUAL UNTIL ALL THE TUBES IN THIS BOX ARE USED UP.

1. PERFORMANCE:

Measuring Range	: 0.5 - 10 ppm (*)	0.25 - 5 ppm
and Sampling Time	: 1.5 minutes	3 minutes



KITAGAWA
KOMYO RIKAGAKU KOGYO

INSTRUCTION MANUAL
CARBON DIOXIDE DETECTOR TUBE

No.126SF

★ READ CAREFULLY THIS INSTRUCTION MANUAL AND THE INSTRUCTIONS OF THE ASPIRATING PUMP PRIOR TO USING THIS PRODUCT.
★ DO NOT DISCARD THIS INSTRUCTION MANUAL UNTIL ALL THE TUBES IN THIS BOX ARE USED UP.

1. PERFORMANCE:

Measuring Range	: 100 - 2000 ppm (*)	200 - 4000 ppm
and Sampling Time	: 2 minutes	1.5 minutes

SENSIDYNE®
Precision Gas Detector Tubes

OPEN

Tube No. 175U

NITROGEN OXIDES

0.5–30ppm

10 Tubes

Lot No.310062

Store in a cool and dark place

nSL17510

Made In Japan

For use with Sensidyne piston-type gas detection pumps only.

CAUTION:

- Leak test pump before use.
- Read enclosed instructions before use.

For technical support call:
Tel: 800.451.9444 / 727.530.3602
Fax: 727.539.0550
E-mail: info@Sensidyne.com

Sensidyne, LP • 1000 112th Circle North, Suite 100
St. Petersburg, FL 33716 USA • www.Sensidyne.com

0.5-5 — 0.5-

Precision gas detector tubes

Kitagawa

NOx

0.5 1 3 5 10 15 ppm

0.5 1 3 5 10 15 mL

0.5 1 3 5 10 15 mL

0062 PAN 150

062 XN 50

062 AN 50

INSTRUCTION MANUAL
NITROGEN OXIDES DETECTOR TUBE
No.175U

KITAGAWA
KOMYO RIKAGAKU KOGYO

★ READ CAREFULLY THIS INSTRUCTION MANUAL AND THE INSTRUCTIONS OF THE ASPIRATING PUMP PRIOR TO USING THIS PRODUCT.
★ DON'T DISCARD THIS INSTRUCTION MANUAL UNTIL ALL THE TUBES IN THIS BOX ARE USED UP.

1. PERFORMANCE:

Measuring Range	1-30ppm	0.5-15ppm (*)
and Sampling Time	: 30 seconds	1 minute

(*) Graduations on the detector tube are based on