phenova™ | A Phenomenex® Company
Certified Reference Materials

ACCREDITED
Proficiency Testing Provider
Certificate # 2427.01

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

July 23, 2025

Mrs. Alejandra Rincon
Corporación Autonoma Regional Para La Defensa De La Meseta de Bucaramanga
Anillo Vial, Giron Lt PR3+958
Giron, Santander 1
Colombia

Dear Alejandra,

Thank you for participating in the R41455 Rapid Return™ Proficiency Testing Study. Enclosed is your final report, which has been carefully reviewed by the PT specialists at Phenova. A final report for your laboratory has been sent to all accrediting agencies you requested at the time of your data submittal.

Thank you again for participating in the R41455 Rapid Return™ Proficiency Testing Study. We appreciate working with you and look forward to our next study. If you have any questions, please call us at 866-942-2978.

phenova™ | A Phenomenex® Company
Certified Reference Materials

ACCREDITED
Proficiency Testing Provider
Certificate # 2427.01

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

Report Definitions:

Assigned Value The Assigned Value is determined from the study mean, gravimetric and volumetric true concentration of an analyte to be analyzed, calculation and/or an appropriate reference value as stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Document (current version), the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation Limits Acceptance Limits are derived from fixed limits, coefficients, constants and calculations stipulated in the EPA National Standards for Water Proficiency Testing Studies Criteria Documents (current version), the National Environmental Laboratory Accreditation Conference Institute (TNI) criteria (ref: TNI FOT tables, TNI PT Committee) and other documents distributed by accrediting agencies as applicable.

Evaluation

Acceptable	The reported value falls within the Acceptance Limits.
Not Acceptable	The reported value falls outside the Acceptance Limits.
No Evaluation	The reported value is non-numeric and can not be evaluated.
NR	As required by the TNI standards and requested by state authorities, any analyte purchased but not reported by your facility is listed as NR (Not Reported).

This report must not be reproduced except in full, without the written approval of Phenova.
Various aspects of Phenova's proficiency testing schemes can from time to time be subcontracted. If subcontracting occurs, it is placed with a competent and qualified subcontractor and Phenova is responsible for all work performed.
This report must not be used to claim product certification, approval, or endorsement by the American Association for Laboratory Accreditation (A2LA) or any agency of the federal government.

phenova™ | A Phenomenex® Company
Certified Reference Materials

ACCREDITED
Proficiency Testing Provider
Certificate # 2427.01

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

State Regulatory Agency Request(s):

No requests were made to send a copy of your R41455 final results to a state agency.

Other Regulatory Agency Request(s):

No requests were made to send a copy of your R41455 final results to an agency.

Third Party Request(s):

No requests were made to send a copy of your R41455 final results to a third party.

phenova™ | A Phenomenex® Company
Certified Reference Materials

ACCREDITED
Proficiency Testing Provider
Certificate # 2427.01

6390 Joyce Drive
100
Golden, CO 80403

Phone 303-940-0033
Fax 866-283-0269
www.phenova.com

Final Report - Rapid Return™

Study: R41455

Opening Date: July 8, 2025 - Closing Date: July 23, 2025

Laboratory: Corporación Autonoma Regional Para La Defensa De La Meseta de Bucaramanga
Anillo Vial, Giron Lt PR3+958
Giron, Santander 1
COLOMBIA

Contact: Mrs. Alejandra Rincon
3104971079

EPA Lab ID:

Dissolved Oxygen (RR-O2-WP) — Lot #: R41455-78

NELAC Code	Analyte	Analysis Date	Analyst	Method Code	Method Description	Units	Assigned Value	Result	Acceptance Limits	Evaluation
Additional State Specific Analytes										
1880	Dissolved Oxygen	7/22/2025	Andres Sarmiento		SM 4500 O- H	mg/L	6.89	7.00	2.76 - 9.65	Acceptable