
AMOZ

CERTIFICATE OF ANALYSIS

Doc. No:	5091AMOZ-CA V20
Batch No:	250552
Code:	5091AMOZ
Exp. Date:	01-2027

This certificate relates to the above specified product batch and number only. The product has been tested according to our standard quality control program. The product is guaranteed when stored according to the instruction in the manual till the indicated expiry date.

Components:	Lot No:	Exp. Date:
Microtiter plate	250510	01-2027
Conjugate	250511	01-2027
Standards 2 - 1 - 0,5 - 0,25 - 0,125 - 0,0625 - 0 ng/ml	250512	03-2028
Standaard 100 ng/ml	250513	03-2028
Sample Dilution Buffer	250462	06-2027
Dilution buffer	250188	03-2027
Rinsing buffer	250333	04-2027
TMB substrate	241194	07-2028
Stop solution	250185	03-2027
Protocol	Code 5091AMOZ [15]02.24	



Validation criteria

Bmax > 0.8 OD.
Blank < 0.2 OD.

Concentration (ng/ml)	Accepted Range %Bmax	Mean %Bmax
STD 1	18 - 32	21
STD 0,5	27 - 49	35
STD 0,25	45 - 65	49
STD 0,125	62 - 78	70

Performed by: TK date: 08-07-2025

Accepted by:

ISO 9001 BUREAU VERITAS Certification

This product is produced under the control of a management system approved by Bureau Veritas certification in accordance with the standard ISO 9001: 2015 Bureau Veritas Certification N°3089368