
Tetracycline

CERTIFICATE OF ANALYSIS

Doc. No:	5091TC-CA V11
Batch No:	250395
Code:	5091TC
Exp. Date:	08-2026

This certificate relates to the above specified product batch and number only. The product has been tested according to our standard quality control program. The product is guaranteed when stored according to the instruction in the manual till the indicated expiry date.

Components:	**Lot No:**	**Exp. Date:**
Microtiter plate	250073	08-2026
Conjugate	250070	08-2026
Standard 2 ng/ml	250113	02-2027
Standard 1000 ng/ml	250114	02-2027
MTC buffer	241042	10-2026
Dilution buffer	250145	02-2027
Rinsing buffer	250058	01-2027
TMB substrate	241133	10-2027
Stop solution	250185	03-2027
Protocol	Code 5091TC [9]02.24	



Validation criteria

Bmax > 0.8 OD.
Blank < 0.2 OD.

Concentration (ng/ml)	Accepted Range %Bmax	Mean %Bmax
STD 1	10 - 17	15
STD 0,5	25 - 42	38
STD 0,25	48 - 70	67
STD 0,125	71 - 88	84

Performed by: BB date: 28-04-2025

Accepted by:
QA/QC Manager

ISO 9001 BUREAU VERITAS Certification

This product is produced under the control of a management system approved by Bureau Veritas certification in accordance with the standard ISO 9001: 2015 Bureau Veritas Certification N°3089368