
Methyl Testosteron

CERTIFICATE OF ANALYSES

Doc. No:	5081MTES-CA V06
Batch No:	250453
Code:	5081MTES
Exp. Date:	08-2026

This certificate relates to the above specified product batch and number only. The product has been tested according to our standard quality control program. The product is guaranteed when stored according to the instruction in the manual till the indicated expiry date.

Components:	**Lot No:**	**Exp. Date:**
Microtiter plate	250300	04-2027
Conjugate	250423	11-2026
Standards 1 - 0,5 - 0,25 - 0,125 - 0,0625 - 0,0313 - 0 ng/ml	250424	11-2026
Antibody	250425	11-2026
Dilution buffer	240866	08-2026
Rinsing buffer	250333	04-2027
TMB substrate	250373	03-2028
Stop solution	250185	03-2027
Protocol	Code 5081MTES [5]02.24	



Validation criteria

Bmax > 0.8 OD.
Blank < 0.2 OD.

Concentration (ng/ml)	Accepted Range %Bmax	Mean %Bmax
STD 0,5	12 - 29	24
STD 0,25	20 - 41	38
STD 0,125	40 - 61	53
STD 0,0625	60 - 80	66

Performed by: SV date: 27-05-2025

Accepted by:
QA/QC Manager

ISO 9001 BUREAU VERITAS Certification

This product is produced under the control of a management system approved by Bureau Veritas certification in accordance with the standard ISO 9001: 2015 Bureau Veritas Certification N°3089368

r-biopharm®