thermoscientific

Certificate of Quality

Product Name: D/M WATER, 5ML, 10/BOX, IVD
Lot Number: 294793
Product Number: T3339-10
Expiration Date: 2027-07-15
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Performance Criteria:

Conductivity	< 50 µS/cm


Physical Characteristics:

Appearance:	Clear, Colorless Broth w/ No Precipitate

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Senior Quality Assurance Engineer