Airgas
an Air Liquide company

Airgas Specialty Gases
Airgas USA, LLC
6141 Easton Road
Bldg 2
Plumsteadville, PA 18949
Airgas.com

CERTIFICATE OF ANALYSIS

Grade of Product: CERTIFIED STANDARD-SPEC

Part Number:	X02NI99C15AJ214	Reference Number:	160-402205436-1A
Cylinder Number:	CC748894	Cylinder Volume:	144.3 CF
Laboratory:	124 - Plumsteadville - PA	Cylinder Pressure:	2015 PSIG
Analysis Date:	Sep 16, 2021	Valve Outlet:	580
Lot Number:	160-402205436-1A		

Expiration Date: Sep 16, 2029

Product composition verified by direct comparison to calibration standards traceable to N.I.S.T. weights and/or N.I.S.T. Gas Mixture reference materials.

ANALYTICAL RESULTS

Component	Req Conc	Actual Concentration (Mole %)	Analytical Uncertainty
OXYGEN	11.00 PPM	10.93 PPM	+/- 5%
NITROGEN	Balance		

Notes: Gross weight: 27.24 kg
Net weight: 4.98 kg

Approved for Release