Certificate of Analysis

Product Name: Dengue Combo Rapid Test Cassette (WB/S/P)

Catalog No.: IDEC-425

Batch No.: DEC25120014

Quantity:132000PCS

Expiry Date:2027-11

Date of Sampling:2025-12-24

Date of Analysis:2025-12-24

Other information:

Buffer Lot: 25125367, Exp:2027-11

Date of manufacture:2025-12-24

QC Item		QC Criterion	QC Result	Conclusion
Physical	Appearance	Good	Good	Pass
Dengue NS1	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
Dengue IgG	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass
Dengue IgM	Positive Sample	Positive	100% Positive	Pass
	Negative Sample	Negative	100% Negative	Pass

Others	/

Final QC Conclusion：	This batch of product met the QC Criteria.

Control

QC supervisor: Eva zhang

Date: 2025.12.24

杭州奥泰生物技术股份有限公司 质检章

Control