
Certificate of Analysis

RIDASCREEN® Chloramphenicol

REF Art. No.	R1511
LOT Lot	22294
Expiry	2026-05-05



Standards

Std. Conc.(ng/L / ng/kg) CV(%)	n mean B/B0
Std1	3
0.00	2.324
1.1	100.0
Std2	3
25.00	1.765
0.8	75.9
Std3	3
50.00	1.402
3.4	60.3
Std4	3
100.00	0.992
2.6	42.7
Std5	3
250.00	0.570
7.2	24.5
Std6	3
750.00	0.276
6.8	11.9

	Lot. No.	Expiry
Microtiter plate	21194	2026-05-05
Standards	22284	2026-07-08
Wash buffer salt Tween	11349	2029-08-31
Conjugate	21284	2026-07-07
Substrate/Chromogen (Red Chromogen Pro)	24084	2026-11-30
Stop solution	21114	2029-03-10

IFU (and other accompanying documents where applicable)

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Please note: The absorbance for the standards may decrease during the shelf life of the kit. The general shape of the curve will remain similar, while the slope might change slightly. Furthermore refer to product leaflet section 8. Indication of instability or deterioration of reagents.

Quality Control Department

Date: 2024-07-17

Quality Management System certified • ISO 9001 • ISO 13485

Remark: This document is created electronically and therefore valid without a signature.