
Certificate of Analysis

RIDASCREEN® Zearalenon ECO

REF Art. No.	R1403
LOT Lot	23265
Expiry	2026-05-18



Standards

Std. Conc.(µg/L / µg/kg) CV(%)	n mean B/B0
Std1	3
0.00	2.065
0.8	100.0
Std2	3
7.50	1.623
0.7	78.6
Std3	3
22.50	1.202
2.4	58.2
Std4	3
67.50	0.812
3.1	39.3
Std5	3
202.50	0.507
3.1	24.6
Std6	3
607.50	0.260
2.7	12.6

	Lot. No.	Expiry
Microtiter plate K	22244	2026-09-10
ECO Extractor	22195	2027-11-05
Standards	21215	2026-05-18
Wash buffer salt Tween	21422	2027-09-30
Conjugate	21215	2026-05-18
Antibody	21215	2026-05-18

	Lot. No.	Expiry
Substrate/Chromogen (Red Chromogen Pro)	25294	2026-11-30
Stop solution	25045	2030-01-23

IFU (and other accompanying documents where applicable)

We herewith certify that this batch has been approved by the Quality Control Department and has met the release criteria.

Please note: The absorbance for the standards may decrease during the shelf life of the kit. The general shape of the curve will remain similar, while the slope might change slightly. Furthermore refer to product leaflet section 8. Indication of instability or deterioration of reagents.

Quality Control Department

Date: 2025-06-25

Quality Management System certified • ISO 9001 • ISO 13485

Remark: This document is created electronically and therefore valid without a signature.