Certificate of Analysis – Certified Reference Material

Certipur® Buffer solution pH 7.00 (25°C)

Certified Reference Material for pH measurement

Producer:	Merck KGaA, Frankfurter Str. 250, 64293 Darmstadt, Germany
Product no.:	1.09407.1000
Lot no.:	HC68096407
Description of CRM:	Certipur® Buffer solution pH 7.00 (25°C) Certified Reference Material for pH measurement
Expiry date:	2029/01/31
Storage:	+15°C to +25°C tightly closed in the original container
Composition:	potassium dihydrogen phosphate / di-sodium hydrogen phosphate

Certified value	Associated uncertainty, $U=k\cdot u$ ($k=2$)
pH value 6.99	**±0.02 (25°C)**

Metrological traceability:	The pH value of this certified buffer solution is directly traceable to primary certified reference materials characterised by PTB and verified by SRMs from NIST. NIST 189c, 188, 185i, 186 Ig, 186 IIg, 187f PTB OX-530/22, TA-582/23, PHT-606/24, PHO-623/25, BO-580/23 *PTB: Physikalisch Technische Bundesanstalt, Braunschweig, Germany* *NIST: National Institute of Standards and Technology, Gaithersburg, USA.*
Measurement method:	pH value is measured with a combined glass electrode after 5-point calibration according to DIN 19268 with reference buffer solutions according to DIN 19266, IUPAC, NIST, Ph.Eur. and USP.
Accreditation:	Merck KGaA, Darmstadt, Germany is accredited by the German accreditation authority DAkkS as registered reference material producer D-RM-15185-01-00 in accordance with ISO 17034.
Certificate issue date:	2026/01/29

ilac-MRA DAkkS Deutsche Akkreditierungsstelle D-RM-15185-01-00

ISO 17034

CRM released by Approving Officer
or delegate of Quality Control

A. Yildirim

Dipl.-Ing. Ayfer Yildirim
Responsible Laboratory Manager

Intended use:	This reference material is intended for use as a calibration standard for pH instruments or pH electrodes or as a control sample for measuring the pH value.
Instructions for handling and correct use:	The pH value is strongly dependent on the temperature. It is therefore necessary to keep the temperature constant within the measurement.
Health and safety information:	Please refer to the Safety Data Sheet for detailed information about the nature of any hazard and appropriate precautions to be taken.
Preparation:	This reference material is prepared gravimetrically from potassium dihydrogen phosphate / disodium hydrogen phosphate and high purity water.

Associated uncertainty:

The expanded uncertainty U_{CRM} is calculated as $U_{CRM}=k \cdot u_{CRM}$, where $k=2$ is the coverage factor for a 95% coverage probability and u_{CRM} is the combined standard uncertainty in accordance to ISO 17034.

The combined uncertainty u_{CRM} is derived from combination of the squared uncertainty contributions:

$$u_{CRM} = \sqrt{u^2_{Characterisation} + u^2_{Homogeneity} + u^2_{Stability}}$$

***u*characterisation:**	is the uncertainty in accordance with DIN EN ISO/IEC 17025 which includes the contributions of the primary reference material and the measuring system. The characterisation measurements have been conducted by our DAkkS accredited calibration laboratory.
***u*homogeneity:**	is the between-bottle variation in accordance with ISO 17034. The assessment of homogeneity is performed by analysis of a representative number of systematically chosen sample units.
***u*stability:**	is the uncertainty obtained from short-term and long-term stability in accordance with ISO 17034. The stability studies are the basis for the quantification of the expiry date of this reference material for the unopened bottle.

Informative values:

Temperature dependence[1]:

Temperature [°C]	Δ pH
5	+ 0.09
10	+ 0.06
15	+ 0.04
20	+ 0.02
25	± 0.00
30	- 0.02
35	- 0.02
40	- 0.03
45	- 0.03
50	- 0.03

[1]Temperature deviation data provided for reference only. Values are not batch-specific and should not be considered certified values.

For more detailed information please read the certification report on our website.

Certificate of analysis revision history:

Certificate version	Date	Reason for version
01	2026/01/29	Initial version

The vibrant M, Supelco, Certipur and Sigma-Aldrich are trademarks of Merck KGaA, Darmstadt, Germany or its affiliates.
Detailed information on trademarks is available via publicly accessible resources.
© 2024 Merck KGaA, Darmstadt, Germany and/or its affiliates. All Rights Reserved.

The life science business of Merck KGaA, Darmstadt, Germany operates as MilliporeSigma in the U.S. and Canada.