



October 6, 2020

Dear valued customer,

Our products undergo extensive testing and validation to ensure our customers receive high quality products. Due to the delicate nature of the products within each TapeStation and Bioanalyzer assay it is not possible to provide all products with a minimum 12-month shelf-life.

Our manufacturing and logistics processes are in place to ensure product integrity is retained when shipped to the customer. Agilent guarantees that products shipped to the customer will have a minimum product stability from delivery date as defined in the table below. The minimum shelf life is guaranteed for direct shipment from Agilent to the end customer; purchase conditions via distributors may differ.

Please find below the list of products for our TapeStation and Bioanalyzer instruments with their guaranteed shelf-life:

TapeStation assays:

Agilent Part No	Agilent Product Description	Product stability from Delivery Date
5067-5582	D1000 ScreenTape	Minimum 6 month
5067-5583	D1000 Reagents	Minimum 6 month
5067-5586	D1000 Ladder	Minimum 6 month
5067-5602	D1000 Sample Buffer	Minimum 6 month
5067-5584	High Sensitivity D1000 ScreenTape	Minimum 6 month
5067-5585	High Sensitivity D1000 Reagents	Minimum 6 month
5067-5587	High Sensitivity D1000 Ladder	Minimum 6 month
5067-5603	High Sensitivity D1000 Sample Buffer	Minimum 6 month
5067-5588	D5000 ScreenTape	Minimum 4 month
5067-5589	D5000 Reagents	Minimum 4 month
5067-5590	D5000 Ladder	Minimum 4 month
5067-5592	High Sensitivity D5000 ScreenTape	Minimum 4 month
5067-5593	High Sensitivity D5000 Reagents	Minimum 4 month
5067-5594	High Sensitivity D5000 Ladder	Minimum 4 month
5067-5365	Genomic DNA ScreenTape	Minimum 4 month
5067-5366	Genomic DNA Reagents	Minimum 4 month
5067-5576	RNA ScreenTape	Minimum 4 month
5067-5577	RNA ScreenTape Sample Buffer	Minimum 4 month
5067-5578	RNA ScreenTape Ladder	Minimum 4 month
5067-5579	High Sensitivity RNA ScreenTape	Minimum 4 month

For Research Use Only. Not for use in diagnostic procedures.



Agilent Part No	Agilent Product Description	Product stability from Delivery Date
5067-5580	High Sensitivity RNA ScreenTape Sample Buffer	Minimum 4 month
5067-5581	High Sensitivity RNA ScreenTape Ladder	Minimum 4 month
5067-5630	Cell-free DNA ScreenTape	Minimum 6 month
5067-5631	Cell-free DNA Reagents	Minimum 6 month
5067-5632	Cell-free DNA Ladder	Minimum 6 month
5067-5633	Cell-free DNA Sample Buffer	Minimum 6 month

Bioanalyzer kits:

Agilent Part No	Agilent Product Description	Product stability from Delivery Date
5067-1504	DNA 1000 kit	Minimum 4 month
5067-1505	DNA 1000 reagents	Minimum 4 month
5067-1506	DNA 7500 kit	Minimum 4 month
5067-1507	DNA 7500 reagents	Minimum 4 month
5067-1508	DNA 12000 kit	Minimum 4 month
5067-1509	DNA 12000 reagents	Minimum 4 month
5067-4626	High Sensitivity DNA kit	Minimum 4 month
5067-4627	High Sensitivity DNA reagents	Minimum 4 month
5067-1515	Protein 80 kit	Minimum 4 month
5067-1516	Protein 80 reagents	Minimum 4 month
5067-1517	Protein 230 kit	Minimum 4 month
5067-1518	Protein 230 reagents	Minimum 4 month
5067-1577	High Sensitivity Protein 250 labeling kit	Minimum 6 month
5067-1575	High Sensitivity Protein 250 kit	Minimum 6 month
5067-1576	High Sensitivity Protein 250 reagents	Minimum 6 month
5067-1578	High Sensitivity Protein 250 ladder	Minimum 6 month
5067-1511	RNA 6000 Nano kit	Minimum 4 month
5067-1512	RNA 6000 Nano reagents	Minimum 4 month
5067-1529	RNA 6000 Nano ladder	Minimum 4 month
5067-1513	RNA 6000 Pico kit	Minimum 4 month
5067-1514	RNA 6000 Pico reagents	Minimum 4 month
5067-1535	RNA 6000 Pico ladder	Minimum 4 month
5067-1548	Small RNA kit	Minimum 4 month
5067-1549	Small RNA reagents	Minimum 4 month
5067-1550	Small RNA ladder	Minimum 4 month

For Research Use Only. Not for use in diagnostic procedures.



Sincerely,

A handwritten signature in black ink, appearing to read "Elisa Viering".

Elisa Viering
Product Manager
Agilent ScreenTape Portfolio
Agilent Technologies

A handwritten signature in black ink, appearing to read "Eva Graf".

Eva Graf
Product Manager
2100 Bioanalyzer system
Agilent Technologies