

Pos Nr	Samplename Remak	Weight	U	Creation Date	State	Height	Hg [ng]	Concentr. [µg/L]	Cal-Factor	Calibration file Date	Method file Date
1 1	BCO	0.1000	ml	06.05.2026 10:49:23	✓	0.0248	0.0000	0.0000	1.0000	<data not saved> 06.05.2026 11:06:23	water method.m80 16.04.2026
2 2	5 ppb	0.1000	ml	06.05.2026 10:50:04	✓	0.0484	0.5000	5.0000	1.0000	<data not saved> 06.05.2026 11:12:19	water method.m80 16.04.2026
3 3	10 ppb	0.1000	ml	06.05.2026 10:50:04	✓	0.0757	1.0000	10.0000	1.0000	<data not saved> 06.05.2026 11:18:16	water method.m80 16.04.2026
4 4	30 ppb	0.1000	ml	06.05.2026 10:50:05	✓	0.1935	3.0000	30.0000	1.0000	<data not saved> 06.05.2026 11:24:12	water method.m80 16.04.2026
5 5	50 ppb	0.1000	ml	06.05.2026 10:50:05	✓	0.3054	5.0000	50.0000	1.0000	<data not saved> 06.05.2026 11:30:09	water method.m80 16.04.2026
6 6	80 ppb	0.1000	ml	06.05.2026 10:50:06	✓	0.4598	8.0000	80.0000	1.0000	<data not saved> 06.05.2026 11:36:07	water method.m80 16.04.2026
7 7	100 ppb	0.1000	ml	06.05.2026 10:50:06	✓	0.6093	10.0000	100.0000	1.0000	<data not saved> 06.05.2026 11:42:04	water method.m80 16.04.2026
8 8	120 ppb	0.1000	ml	06.05.2026 10:50:07	✓	0.6473	12.0000	120.0000	1.0000	<data not saved> 06.05.2026 11:48:02	water method.m80 16.04.2026
9 9	150 ppb	0.1000	ml	06.05.2026 10:50:07	✓	0.7666	15.0000	150.0000	1.0000	<data not saved> 06.05.2026 11:53:59	water method.m80 16.04.2026
10 10	STD 30 ppb	0.1000	ml	06.05.2026 11:54:42	✓	0.2064	3.1976	31.9764	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
11 11	STD 100 ppb	0.1000	ml	06.05.2026 11:55:13	✓	0.5543	9.9882	99.8823	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
12 12	2026-1430-W	0.1000	ml	06.05.2026 12:19:15	✓	0.0304	0.1815	1.8147	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
13 13	2026-1431-W	0.1000	ml	06.05.2026 12:19:16	✓	0.0169	0.0000	0.0001	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
14 14	2026-1432-W	0.1000	ml	06.05.2026 12:19:17	✓	0.0131	0.0000	0.0001	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
15 15	2026-1433-W	0.1000	ml	06.05.2026 12:19:18	✓	0.0133	0.0000	0.0001	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
16 16	2026-1434-W	0.1000	ml	06.05.2026 12:19:19	✓	0.0115	0.0000	0.0001	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
17 17	2026-1435-W	0.1000	ml	06.05.2026 12:19:19	✓	0.0122	0.0000	0.0001	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
18 18	2026-1436-W	0.1000	ml	06.05.2026 12:19:20	✓	0.0107	0.0000	0.0001	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
19 19	2026-1437-W	0.1000	ml	06.05.2026 12:19:20	✓	0.0093	0.0000	0.0001	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
20 20	MFL 2026-1437-W	0.1000	ml	06.05.2026 14:16:48	✓	0.2572	4.1131	41.1315	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
21 21	MFLD 2026-1437-W	0.1000	ml	06.05.2026 14:16:48	✓	0.3085	5.0607	50.6070	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
22 22	BFL 60 ppb	0.1000	ml	06.05.2026 14:19:39	✓	0.3686	6.2033	62.0326	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	water method.m80 16.04.2026
23 -	manual BV	0.0000	ml	06.05.2026 14:42:37	✗ B	0.0037	0.0000	0.0000	1.0000	Curva-2026-05-06-S.c80 06.05.2026 11:54:34	16.04.2026

