

Pos Nr	Samplename Remak	Weight	U	Creation Date	State	Height	Hg [ng]	Concentr. [µg/L]	Cal-Factor	Calibration file Date	Method file Date
1 1	BCO	0.1000	ml	14.05.2026 10:41:06	✓	0.0210	0.0157	0.1571	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
2 2	STD 30 ppb	0.1000	ml	14.05.2026 10:41:15	✓	0.1914	2.9852	29.8519	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
3 3	STD 100 ppb	0.1000	ml	14.05.2026 10:41:16	✓	0.5614	10.4205	104.2051	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
4 4	2026-1465	0.1000	ml	14.05.2026 11:03:14	✓	0.0259	0.0976	0.9758	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
5 5	2026-1466	0.1000	ml	14.05.2026 11:03:38	✓	0.0134	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
6 6	2026-1474	0.1000	ml	14.05.2026 11:03:40	✓	0.0171	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
7 7	2026-1475	0.1000	ml	14.05.2026 11:04:23	✓	0.0093	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
8 8	2026-1476	0.1000	ml	14.05.2026 11:04:24	✓	0.0122	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
9 9	2026-1477	0.1000	ml	14.05.2026 11:04:29	✓	0.0092	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
10 10	2026-1478	0.1000	ml	14.05.2026 11:04:30	✓	0.0218	0.0291	0.2915	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
11 11	2026-1483	0.1000	ml	14.05.2026 11:04:33	✓	0.0079	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
12 12	2026-1484	0.1000	ml	14.05.2026 11:04:47	✓	0.0171	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
13 13	2026-1485	0.1000	ml	14.05.2026 11:04:48	✓	0.0120	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
14 14	BFL 60 ppb	0.1000	ml	14.05.2026 13:57:37	✓	0.3382	5.7443	57.4433	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
15 15	2026-1486	0.1000	ml	14.05.2026 14:10:03	✓	0.0305	0.1750	1.7504	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
16 16	2026-1487	0.1000	ml	14.05.2026 14:10:19	✓	0.0196	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
17 17	2026-1488	0.1000	ml	14.05.2026 14:10:20	✓	0.0162	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
18 18	2026-1489	0.1000	ml	14.05.2026 14:10:21	✓	0.0702	0.8498	8.4982	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
19 19	2026-1490	0.1000	ml	14.05.2026 14:10:24	✓	0.0221	0.0342	0.3419	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
20 20	2026-1491	0.1000	ml	14.05.2026 14:10:25	✓	0.0243	0.0712	0.7116	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
21 21	2026-1492	0.1000	ml	14.05.2026 14:10:25	✓	0.0184	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
22 22	2026-1493	0.1000	ml	14.05.2026 14:10:26	✓	0.0171	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
23 23	2026-1494	0.1000	ml	14.05.2026 14:10:26	✓	0.0124	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
24 24	2026-1495	0.1000	ml	14.05.2026 14:10:27	✓	0.0129	0.0000	0.0001	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
25 25	MFL 2026-1492	0.1000	ml	14.05.2026 15:33:00	✓	0.2714	4.4638	44.6378	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026



Pos Nr	Samplename Remak	Weight	U	Creation Date	State	Height	Hg [ng]	Concentr. [µg/L]	Cal-Factor	Calibration file Date	Method file Date
26 26	MFLD 2026-1493	0.1000	ml	14.05.2026 15:33:15	✓	0.3076	5.1511	51.5108	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
27 27	BFL 60 ppb	0.1000	ml	14.05.2026 15:58:01	✓	0.3284	5.5530	55.5297	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	water method.m80 16.04.2026
28 -	Burn out amalgamator	0.0000	ml	20.05.2026 09:14:32	✓ B	0.1484	2.2135	0.0000	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	16.04.2026
29 -	manual BV	0.0000	ml	20.05.2026 09:15:32	✓ B	0.0046	0.0000	0.0000	1,0000	Curva 2026-05-13-W.c80 13.05.2026 11:36:20	16.04.2026

