

ANALYSIS CERTIFICATE

CODE: 181671

BATCH: 0002547029

PRODUCT: EDTA Disodium Salt 0.01 mol/l (0.01M) volumetric solution

ISSUE DATE: 08/07/2024

RETEST DATE: 07/2028

SPECIFICATIONS	GUARANTEE VALUE	ACTUAL VALUE
Titer at 20°C	0,999 - 1,001	1,000
Uncertainty	± 0,0015	
Determination method	C	

Balance

Calibrated periodically against traceable weights.

Volumetric glassware

Class A according NBS, document NBSIR-74661-Dec 1974

Determination of concentration

Standardization with Zinc Oxide standard, traceable itself with primary standard NIST, SRM: 682

Indicator: Eriochrome Black T

DETERMINATION METHODS	SUBSTANCE	SRM/NIST
A Argentimetric titration	Potassium Chloride	999b
B Permanganometric titration	Sodium Oxalate	8040
I Redox titration	Potassium Dichromate	136f
C Complexometric titration	Zinc	682
D Iodometric titration	Arsenic III Oxide	3103a
E Acidimetry	Tris (Hydroxymethyl) Aminomethane	723e
H Alkalimetry	Potassium Hydrogen Phthalate	84l
F Precipitation titrimetry	Potassium Chloride	999b
G Non aqueous acid titration	Potassium Hydrogen Phthalate	84l
J Non aqueous alkali titration	Benzoic Acid	350b
K Acidimetry	Sodium Carbonate	351a

Notes

- ☐ All chemicals used for preparing this solution are for analysis, ACS quality.
- ☐ The stability of prepared solution must be checked periodically by the user.
- ☐ Don't use the solution if it is contaminated accidentally.

Panreac Química SLU

C/Garraf, 2
Polígono Pla de la Bruguera
E-08211 Castellar del Vallès
(Barcelona) Spain
Phone +34 937 489 400
info.es@itwreagents.com
www.itwreagents.com



QA Director
T. Llauradó

CODE: 181671

BATCH: 0002547029