

Certificate of Analysis

Product name: Zero Oxygen Solution Set (2 components)
Product code: HI7040L
Lot number: S0164/25
Best use before: November 2030 or one month after mixing
Date of analysis: 2025/12/05
Certified value, % O₂: 0.0 ± 0.1 @ 25 °C
Method of standardization:

This quality product is standardized by Iodometric Methods, according to Standards Methods for the Examination of Water and Waste water, 20th ed. 1998, p. 4-129, 4500-O B. Standard solutions used are prepared with deionized water for analytical use ISO 3696 / BS 3978.

Balances and thermometers used are checked with certified reference materials.

Uncertainty U:

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

Components identification:

Component code:	Lot number:	Expiration Date:
HI7040-1	2498	November 2030
HI7040-2	2591	December 2030

Reference number: 05L59

QA manager: Eugenia Tulbure 

CERT004M0016