3 Mansfield Road
Western Avenue Business Park
London W3 0BZ - UK
Tel. +44 208 993-4669
Fax +44 208 993-2996
www.fc2femalecondom.com

FICHA TÉCNICA

FC2 CONDÓN FEMENINO

A- DESCRIPCIÓN DEL DISPOSITIVO:

Condón suave y flexible que cubre la vagina. El FC2 está elaborado de Polímero de Nitrilo, que es un material sintético no alergénico. Tiene un anillo interno que se utiliza para su inserción y que sostiene el condón en su lugar, y un anillo externo que yace tendido sobre los labios durante el coito. El dispositivo viene en un solo tamaño y está diseñado para utilizarse una única vez. El condón está pre-lubricado con un lubricante hecho a base de silicona.

El anillo interno se utiliza para insertar el condón. Se toma este anillo entre el pulgar y el índice, apretando hasta que se junte y se forme una punta, la cual se inserta en la vagina. Una vez el anillo esté en la vagina, se desliza el dedo índice o el dedo medio dentro del condón y empuja el anillo interior hasta donde sea posible.

El anillo externo y una pequeña parte del condón permanecen por fuera de la vagina. El pene del compañero debe ser guiado hacía el condón y el anillo externo permanecerá tendido sobre la vulva durante el coito. Una vez retirado el pene, el anillo externo se retorcerá para retener el contenido eyaculado y el condón es extraído de la vagina. El dispositivo debe ser desechado en un basurero y no en el inodoro.

El condón femenino reviste la vagina y el cuello del útero y se adapta a todas las formas y tamaños de órganos sexuales masculinos y femeninos. Retiene el semen después de la eyaculación, con lo cual actúa como método anticonceptivo y evita un embarazo no planificado, actúa igualmente como una barrera contra virus y bacterias que causan las infecciones de transmisión sexual, incluido el VIH.

El condón femenino también ofrece protección adicional contra las infecciones de transmisión sexual por contacto, como el HPV, al recubrir los órganos sexuales externos femeninos y la base del pene.

Contiene un lubricante no-espermicida, a base de silicona. Puede añadirse más lubricante a base de agua o aceite, ya sea en el interior o el exterior del condón o directamente en el pene.

B- COMPONENTES:

Componentes	Compuestos
Funda	Nitrilo (41%)
Anillo externo (en el extremo abierto del condón)	Nitrilo (41%)
Anillo interno	Poliuretano (24%)
Lubricante	Silicona (15%)
Intervalo de dosis	1,00 – 2,00 g
Envase	Poliéster/polietileno (20%)

C- DIMENSIONES:

Dimensiones	Especificación
Longitud	163 – 190 (mm)
Anchura (25%)	76 – 90 (mm)
Anchura (50%)	76 – 90 (mm)
Anchura (75%)	76 – 90 (mm)
Grosor (25%)	65 – 105 (µ)
Grosor (50%)	65 – 95 (µ)
Grosor (75%)	60 – 95 (µ)
Diámetro del anillo superior	67 mm mínimo
Grosor del anillo superior	2,9 – 3,6 mm
Diámetro externo del anillo interno	56,5 – 58,5 mm
Grosor del anillo interno	3,3 – 3,6 mm
Altura del anillo interno	4,60 – 5,10 mm

D- PROPIEDADES DE ROTURA:

Longitud total del dispositivo a someter a prueba de rotura.

Límite inferior del volumen de rotura	5,0 litros
Límite inferior de la presión de rotura	3,45 k Pa

E- PÉRDIDAS DE AGUA:

0 – 25 mm	Sin agujeros visibles
25 mm – extremo cerrado	AQL 0,25

F- VIDA ÚTIL:

Cinco años determinados por medio de un estudio en tiempo real a temperatura y humedad elevadas.

G- CONDICIONES DE ALMACENAMIENTO:

Se recomienda “almacenar en lugar seco y protegido de la luz del sol directa”.

H- NORMATIVAS APLICABLES:

The Female Health Company (UK) Plc utiliza las siguientes directrices y normas reglamentarias:

- ISO 9001:2015 Sistemas de gestión de calidad
- ISO 10993:2010 Evaluación Biológica de Productos Sanitarios
- EN ISO 13485:2016 Productos Sanitarios – Sistemas de gestión de calidad
- EN ISO 14971:2012 Productos Sanitarios – Aplicación de la gestión de riesgos
- ISO 2859:2006 Procedimientos de Muestreo
- EN 980:2008 Símbolos Gráficos de Productos Sanitarios
- ISO 25841:2017 Condones Femeninos – Requisitos y métodos de sometimiento a pruebas
- Norma CFR de la FDA estadounidense, Título 21, Partes 800-1299

- Directiva CE de Productos Sanitarios 93/42/CEE
- Directiva CE de Productos Sanitarios 2007/47/CE
- Norma canadiense sobre Dispositivos Médicos SOR/98-282
- Preservativos Femeninos: Especificaciones generales, precalificación y directrices de adquisición, 2012 UNFPA

La ISO 10993:2010, incluye pruebas de biocompatibilidad, citotoxicidad (acción destructiva en determinadas células), mutagenicidad (capacidad de inducir mutaciones en las células), sensibilización, irritación y toxicidad sistémica (posibles efectos adversos en los órganos y tejidos del cuerpo).

I- POTENCIALES USUARIOS:

Mujeres y hombres sexualmente activos pueden utilizar el FC2. Además, pueden utilizarlo:

- Hombres y mujeres sensibles al látex
- Mujeres que estén menstruando
- Mujeres que estén embarazadas
- Mujeres que hayan dado a luz recientemente
- Mujeres que estén en la (pre o post) menopausia
- Mujeres a las que les hayan practicado una histerectomía

J- BENEFICIOS:

- El FC2 es un **método de doble protección**, es decir que protege tanto contra un embarazo no planificado como contra las infecciones de transmisión sexual (ITS), incluido el VIH.
- El FC2 se puede insertar antes de iniciar las relaciones sexuales, con lo cual no es necesario interrumpir la espontaneidad sexual para ponerse un condón.
- El FC2 se adapta perfectamente a la vagina y permite que el pene se mueva libremente dentro del condón durante las relaciones sexuales.
- El FC2 se calienta rápido hasta alcanzar la temperatura corporal, por lo tanto, produce una sensación cálida y natural.
- El FC2 está lubricado, pero se puede añadir más lubricante a base de agua o aceite en el interior o en el exterior del condón o directamente en el pene.
- El FC2 es un producto no alergénico y una excelente opción para hombres y mujeres que tienen alergia al látex.
- El pene no tiene por qué estar erecto para usar el FC2 y no es necesario retirarlo inmediatamente después de la relación sexual.

K- CONTRAINDICACIONES, ADVERTENCIAS/PRECAUCIONES:

El FC2 es un dispositivo libre de hormonas y por lo tanto no causa efectos secundarios.

Al igual que cualquier otro condón, el FC2 se debe insertar con cuidado. No se debe abrir el paquete con tijeras, cuchillos o con los dientes y tener especial precaución al manipular el condón con las uñas largas.

Se han notificado menos de un 1% de roturas en el FC2. Si se rompe el FC2, es necesario quitarlo e inmediatamente inserir un condón nuevo.

L- MÉTODOS DE DESECHO

Después de quitar el condón suavemente, se recomienda envolverlo en un pañuelo de papel o en el paquete vacío y tíralo a la basura.

M- USO COMBINADO CON OTROS ANTICONCEPTIVOS:

El FC2 se puede usar con la píldora, inyecciones, dispositivo intrauterino (DIU), implantes, postesterilización y postvasectomía para proteger contra las ITS, incluido el VIH. El FC2 no se puede usar con el diafragma ni con el NuvaRing, pues el anillo interior del FC2 se aloja en el mismo espacio que dichos dispositivos.

TO WHOM IT MAY CONCERN

As a responsible authorized representative of **The Female Health Company (UK) Plc**. I declare that all the tests detailed in the document of Verification and Validation of Design for the FC2 product, are all the tests that are made to the product during its manufacturing process.

Consider that these tests were performed based on ISO25841.

I hereby certify that the above statements are true and correct to the best of my knowledge.

This document is issued for the reference of the Authorities of the following countries: Argentina, Bolivia, Chile, Colombia, Ecuador, Paraguay, and Uruguay.

Signed by

Samina Naje 

Sr.RA/QA Manager

The Female Health Company (UK) Plc

Date 18.11.25

TO WHOM IT MAY CONCERN

VERIFICATION AND VALIDATION OF DESIGN

PRODUCT FC2

Device description

It is a soft and flexible condom that covers the vagina, and is made of nitrile polymer, which is a non-allergic synthetic material. It has an inner ring that is used for insertion and that holds the condom in place, and an outer ring that lies on the lips during intercourse. The condom comes in a single size and is designed to be used only once. The condom is pre-lubricated with a lubricant made from silicone.

The ring is used to insert the condom. Take this ring between the thumb and forefinger, squeezing until it comes together and forms a tip, which is inserted into the vagina. Once the ring is in the vagina, slide the index finger or middle finger into the condom and push the inner ring as far as possible.

The outer ring and a small part of the condom remain outside the vagina. The partner's penis must be guided to the condom and the outer ring will remain lying on the vulva during intercourse. Once the penis is removed, the outer ring will be twisted to retain the ejaculated contents and the condom is extracted from the vagina. the device must be disposed of in a dumpster and not in the toilet.

The female condom lines the vagina and cervix and adapts to all male and female sexual body shapes and sizes. It retains semen after ejaculation, which acts as a contraceptive method and prevents unplanned pregnancy, also acts as a barrier against viruses and bacteria that cause sexually transmitted infections, including HIV.

The female condom also offers additional protection against sexually transmitted infections by contact, such as HPV, by coating the female external sex organs and the base of the penis.

Contains a non-spermicidal lubricant, based on silicone. More water-based lubricant or oil can be added, either inside or outside the condom or directly on the penis.

Indications of use

FC2 is a condom that helps in family planning and in the prevention of and infections of sexual transmission. The use of condoms is the safest non-medicated contraceptive method known, in addition to the contraceptive action it prevents sexually transmitted infections such as AIDS, syphilis, gonorrhea, chlamydia and herpes.

Verification and validation

The tests were performed to demonstrate performance as intended.

Female Health Company Lot Number		F7035
Expiry Date		2030/01
Quantity		130,000
Components:	**Compound**	
Sheath	Nitrile	Conforms
Outer Ring	Nitrile	Conforms
Inner Ring	Polyurethane	Conforms
Lubricant	Silicone	Conforms
Sachet	Polyester/Polyethylene	Conforms
Dimensions:	**Specification**	
Length	168 - 188(mm)	177
Width (25%)	79 - 83(mm)	81
Width (50%)	79 - 83(mm)	81
Width (75%)	79 - 83(mm)	82
Thickness (25%)	81-109(µ)	97
Thickness (50%)	74-100(µ)	87
Thickness (75%)	70-94(µ)	80
Top Ring Diameter	67 mm minimum	72
Top Ring Thickness	2.9 - 3.6 mm	3.1
Inner Ring Height	4.60 - 5.10 mm	4.99
Inner Ring Thickness	3.3 - 3.6 mm	3.4
Inner Ring External Diameter	56.5 - 58.5 mm	57.5
Physical Properties:		
Complete Length of Device To Be Tested For Burst Test		
Burst Pressure	3.45(kPa) minimum	6.77
Burst Volume	5.0 litres minimum	16.6
Number of Burst Defects		0
Lubricant Dosage	1.00 - 2.00 g	1.67
Sachet Integrity	AQL 2.5	0 /32
Inner ring hardness Shore A	77 - 97	86
Visible Defects		
Critical	AQL 0.4	0
Non Critical	AQL 2.5	2
Water Leakage:		
0 - 25mm	Exclusion Zone	0/315
Non Visible Holes 25mm - closed end/		
Visible holes complete device	AQL 0.25	0/315

Risk analysis:

The process and design of Risk Analysis was carried out in accordance with ISO 14971 to establish the risks associated with medical devices and risk mitigation.

Biocompatibility

FC2 complies with the requirements established by 10993: 2010, Biological evaluation of medical devices. Part 1: Evaluation and trials through a risk management process.

Female condoms are classified as surface devices (mucosal membranes) with the limit of contact with the body less than 24 hours. The following tests are required by ISO 10993-1:

1. Cytotoxicity
2. Sensitization
3. Irritation or intracutaneous Reactivity (Irritation)

The results of the biocompatibility tests are summarized in the following table:

TEST	RESULT	REFERENCE
CYTOTOXICITY	DOES NOT PRODUCE CYTOTOXICITY	ISO10993-5
SKIN SENSITIZATION (POLAR+NON-POLAR EXTRACTS)	DOES NOT CAUSE SKIN SENSITIZATION	ISO10993-10
VAGINAL IRRITATION (SALINE EXTRACT)	DOET NOT CUASE IRRITATION	ISO10993-10

Medicinal substances incorporated within the device

Female Prudence does not contain any substance that can be considered as a medicinal substance.

Stability

A real-time stability study was carried out to demonstrate a useful life of 5 years, under the following conditions: 30 ° C and 65% RH.

Sterilization

The product is not sterile

Clinical Evidence

A clinical study was carried out to demonstrate the acceptability of the use of female condoms. See study in the Clinical Study section.

TO WHOM IT MAY CONCERN



Samina Naje

Sr.RA/QA Manager

The Female Health Company (UK) Plc

Date 18.11.25

Current issue date: 11 July 2025
Expiry date: 10 July 2028
Certificate identity number: 10694044

Original approval(s):
ISO 14001 - 11 July 2016

Certificate of Approval

This is to certify that the Management System of:

The Female Health Company (M) Sdn. Bhd. (772707-P)

No. 1A, Jalan CJ 1/4, Kawasan Perindustrian Cheras Jaya, 43200 Balakong, Selangor, Malaysia

has been approved by LRQA to the following standards:

ISO 14001:2015

Approval number(s): ISO 14001 – 0049214

The scope of this approval is applicable to:

Assembly and in-process testing of nitrile female condoms including in process testing but excluding final product testing and release.

Luis Cunha

Area Operations Manager - North Asia & SAMEA

Issued by: LRQA Limited

UKAS MANAGEMENT SYSTEMS
0001

LRQA Group Limited, its affiliates and subsidiaries and their respective officers, employees or agents are, individually and collectively, referred to in this clause as 'LRQA'. LRQA assumes no responsibility and shall not be liable to any person for any loss, damage or expense caused by reliance on the information or advice in this document or howsoever provided, unless that person has signed a contract with the relevant LRQA entity for the provision of this information or advice and in that case any responsibility or liability is exclusively on the terms and conditions set out in that contract.
Issued by: LRQA Limited, 1 Trinity Park, Bickenhill Lane, Birmingham B37 7ES, United Kingdom

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Deutsche Akkreditierungsstelle D-ZM-11321-01-00

Product Service

Certificate

No. Q5 103602 0001 Rev. 03

Holder of Certificate: **The Female Health Company (UK) Plc**
Unit 23 Park Royal Metro Centre
Britannia Way
Park Royal
London, NW10 7PA
UNITED KINGDOM

Certification Mark:



Scope of Certificate: **Design, Development, Manufacture and Distribution of Female Condom**

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:Q5 103602 0001 Rev. 03

Report No.: 713382081_Q5

Valid from: 2025-10-08
Valid until: 2028-03-14

Date, 2025-10-08

Christoph Dicks
Head of Certification/Notified Body

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Deutsche Akkreditierungsstelle D-ZM-11321-01-00

Product Service

Certificate

No. Q5 103602 0001 Rev. 03

Applied Standard(s):

ISO 13485:2016
(EN ISO 13485:2016/AC:2018, EN ISO 13485:2016/A11:2021)
Medical devices - Quality management systems - Requirements for regulatory purposes

Facility(ies):

The Female Health Company (UK) Plc
Unit 23 Park Royal Metro Centre, Britannia Way, Park Royal, London, NW10 7PA, UNITED KINGDOM

Design, Development, Manufacture and Distribution of Female Condoms.

The Female Health Company (M) Sdn Bhd
No 1A, Jalan CJ 1/4, Kawasan Perindustrian Cheras Jaya, 43200 Balakong, Selangor, MALAYSIA

Manufacture and Distribution of Female Condoms

"-"

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Benannt durch/Designated by
Zentralstelle der Länder für Gesundheitsschutz bei Arzneimitteln und Medizinprodukten
www.zlg.de
BS-MDR-099

Product Service

EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III (Class IIa and Class IIb Devices)

No. G10 103602 0002 Rev. 02

Manufacturer:

The Female Health Company (UK) Plc
Unit 23 Park Royal Metro Centre
Britannia Way
Park Royal
London, NW10 7PA
UNITED KINGDOM

SRN Manufacturer - GB-MF-000029355

Authorized Representative:

Advena Limited
Tower Business Centre, 2nd Flr, Tower Street, Swatar, BKR 4013, MALTA

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result.
The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis.
The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment shall also include an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.
For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G10 103602 0002 Rev. 02

Report No.: 713382081

Preceding Certificate No.: G10 103602 0002 Rev. 01

Valid from: 2025-10-16
Valid until: 2027-07-20

Date of Initial Issuance: 2022-07-21

Christoph Dicks
Head of Certification/Notified Body

Issue date: 2025-10-16

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123
TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Benannt durch/Designated by
Zentralstelle der Länder für Gesundheitsschutz bei Arzneimitteln und Medizinprodukten
www.zlg.de
BS-MDR-099

TÜV SÜD Product Service

EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 103602 0002 Rev. 02

Classification:	Class IIb
Device Group:	U110203 - FEMALE CONDOMS
Intended Purpose:	A single use, non-sterile barrier contraceptive device and to help prevent the transmission of sexually transmitted infections. For vaginal use only.
The validity of this certificate depends on conditions and/or is limited to the following:	./.

Revision History:

Rev.	Dated	Report	Description	
00	2022-07-21	75950573	-	
01	2024-05-16	75959030	Amended: Other	SRN Manufacturer added
02	2025-10-16	713382081	Amended: Change of certificate holder's data	

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123
TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

TÜV®

This leaflet explains how to use the FC2 FEMALE CONDOM® contraceptive device and answers some questions women often ask about its use.

Please read this leaflet carefully before you use the FC2 FEMALE CONDOM for the first time.

FC2 FEMALE CONDOM

The FC2 Female Condom is indicated for preventing pregnancy, HIV/AIDS, and other sexually transmitted infections (STIs). The FC2 Female Condom is a single-use internal condom.

IMPORTANT INFORMATION

Natural rubber latex condoms for men are highly effective at preventing sexually transmitted infections, including HIV/AIDS, if used correctly.

If you are not going to use a male natural rubber latex condom, you can use the FC2 Female Condom to help protect yourself and your partner.

The FC2 Female Condom only works when you use it. Use it correctly every time you have sex.

Before you try the FC2 Female Condom, be sure to read the directions and learn how to use it correctly.

Manufactured for:
Veru Inc.

2916 N. Miami Avenue, Suite 1000
Miami, FL 33127
www.fc2.us.com

FC2 Female Condom is a registered trademark of Veru Inc.

This product is protected by U.S. 7,726,316 B1, Pat. 8,100,128 & others.
G0073 Date: November 2021

READ THIS BEFORE USING THE FC2 FEMALE CONDOM

The FC2 Female Condom is indicated for preventing pregnancy, HIV/AIDS, and other sexually transmitted infections (STIs). The FC2 Female Condom is a single-use internal condom. If you are being treated for a sexually transmitted infection (STI) or think you have one, see your healthcare provider.

IMPORTANT THINGS TO KNOW

Important things to note for the best protection with the FC2 Female Condom

1. Before you try the FC2 be sure to read the directions and learn how to use it correctly.
2. Practice putting the FC2 in the vagina without having sex to get familiar with how it works.
3. Make sure the FC2 is in place before the penis enters.
4. No studies have been conducted to demonstrate if a spermicide is compatible with FC2 material or if using a spermicide increases the FC2's effectiveness.
5. Store the FC2 at normal room temperature.
6. The following are problems that have been reported when using the FC2 during sex:
   - the penis can push the FC2 condom (including outer ring) inside the vagina;
   - the penis slips to the side of the FC2 during intercourse and enters the vagina without FC2 protection;
   - the FC2 Female Condom breaks during use; and
   - the FC2 slips completely out of the vagina.

   In each of the scenarios described above, the risk of vaginal exposure to semen is increased. If any of these occur, you should dispose of that female condom and use a new one.
7. No studies have been conducted to determine the FC2's effectiveness in preventing HIV/AIDS and other sexually transmitted infections (STIs) when used anally.

USING THE FC2

- The FC2 Female Condom is inserted before sex. Most women insert the FC2 Female Condom between 2 to 20 minutes before sex.
- The FC2 Female Condom should be removed after sex and before you stand up. It is for one-time use.
- The FC2 can be used with most commercially available water, silicone, and mineral based sexual lubricants.

BENEFITS OF USING THE FC2

- Using the FC2 during sexual intercourse can help protect against getting a sexually transmitted infection or getting pregnant.
- The FC2 does not contain natural rubber latex.
- The FC2 provides women with an option for protection which they can initiate.
- Women can insert the FC2 ahead of the sex act. This allows for a more spontaneous act because an erection is not needed to insert the FC2.

RISKS OF USING THE FC2

- Use the FC2 Female Condom correctly every time you have sex. If the FC2 is not used correctly every time, your risk of getting a sexually transmitted infection or becoming pregnant will be higher.
- Incorrect use may result in the FC2 not providing protection. To avoid risk of incorrect use, read the instructions carefully prior to use.
- Research has shown that the following events are rare but can occur when using the FC2:
  - discomfort during insertion
  - pain after insertion, before sex
  - discomfort during sex
  - burning sensation, rash or itching.

SOME FEATURES OF USING THE FC2

- You can insert the FC2 Female Condom yourself. It gives you a way to protect yourself.
- The FC2 Female Condom warms up as soon as you insert it. It is both strong and soft.

DESCRIPTION

The FC2 Female Condom is a soft, loose-fitting sheath. It has a ring at each end. The inner ring at the closed end is made of polyurethane and is used to help insert the FC2 Female Condom inside the vagina and helps to hold it in place. It is not connected to the female condom. The outer rim of the sheath, on the open end stays outside the vagina. The FC2 Female Condom comes pre-lubricated with silicone.



ABOUT THE FC'S EFFECTIVENESS

The effectiveness of the FC2 Female Condom is based on a study conducted in South Africa comparing the FC2 Female Condom to the earlier version FC1 Female Condom[1]. This comparative study evaluated the FC2 and FC1 with respect to the four possible problems described earlier in the labeling:

- the female condom may break during use;
- the female condom may slip out of the vagina during use;
- the penis may push the female condom into the vagina; and
- the penis may push past the condom into the vagina (unprotected).

These four problems are the only ways that the vagina would be exposed to semen when using the FC2 Female Condom and this comparative study showed that the rates of these problems with the FC1 and FC2 female condoms are similar. Although the FC2 was not studied for contraception or STI protection, the FC2 effectiveness can be inferred on the basis of this comparative study and an earlier study that demonstrated the effectiveness of the FC1 for contraception. These results are given in the table below.

Below is information on how well different contraceptives protect you against pregnancy. The numbers are based upon typical use and perfect use. Typical use rates indicate that the method may have been used incorrectly, may have not been used with every sex act, or may have failed during use. Perfect use rates indicate that the method was used correctly with every sex act but failed anyway.

% OF WOMEN EXPERIENCING AN UNINTENDED PREGNANCY WITHIN THE FIRST YEAR OF USE

CONTRACEPTION METHOD	TYPICAL USE	PERFECT USE
No method	85	85
Spermicides	29	18
Withdrawal	27	4
Fertility awareness-based methods	25	–
Barrier Methods		
Cervical Cap		
No previous births	16	9
Previous births	32	20
Diaphragm + Spermicide	16	6
Condom-		
FC1 Female Condom	21	5
Male Condom	15	2
Hormonal Methods		
Combined pill and progestin-only pill	8	0.3
Evra Patch®	8	0.3
NuvaRing®	8	0.3
Depo-Provera®	3	0.3
Implanon®	0.05	0.05
IUD		
ParaGard® (copper T)	0.8	0.6
Mirena® (LNG-IUS)	0.2	0.2
Sterilization		
Female Sterilization	0.5	0.5
Male Sterilization	0.15	0.10

Emergency Contraceptive Pills: Treatment initiated within 72 hours after unprotected intercourse reduces the risk of pregnancy by at least 75%.

Lactational Amenorrhea Method: LAM is a highly effective, *temporary* method of contraception.

Source: Trussell J. Contraceptive efficacy. In Hatcher RA, Trussell J, Nelson AL, Cates W, Stewart FH, Kowal D. Contraceptive Technology: Nineteenth Revised Edition. New York NY: Ardent Media, 2007; www.contraceptivetechnology.org

1 Reference: M. Beksinska et al. Performance of the Reality polyurethane female condom and a synthetic latex prototype: a random crossover trial among South African women. Contraception. 2006 Apr; 73(4):386-93.

HOW TO USE THE FC2 FEMALE CONDOM

1 TO OPEN THE PACKET



Fig. A
Packet holds one FC2 Female Condom

- Locate the arrow at the tear notch on the upper right-hand corner and tear open (Fig. A).
- Rub the outside of the pouch together to be sure the lubrication is evenly spread inside the pouch from the bottom to the top.
- Take out the FC2 and look at it closely to make sure it is not damaged.

2 TO INSERT THE FC2 FEMALE CONDOM




Fig. B

- Find a comfortable position. Try standing with one foot up on a chair, or sit with knees apart, or squat down.
- Be sure the inner ring is at the bottom, closed-end of the pouch (Fig. B).

3 NEXT STEPS

- Hold the pouch with the open end hanging down. While holding the outside of the pouch, squeeze the inner ring with your thumb and middle finger.
- Place your index finger between the thumb and middle finger and keep squeezing the inner ring (Fig. C (a)). Or just squeeze – whichever is easier for you (Fig. C (b)).
- Still squeezing the FC2 Female Condom with your other hand, spread the lips of the vagina as if you were getting ready to insert a tampon.
- Insert the squeezed FC2 Female Condom as shown in Fig. D.




Fig. C


Fig. D

Take your time. If the FC2 Female Condom is slippery to insert, let it go and start over.

4 NEXT STEPS.

- Now push the inner ring and the pouch the rest of the way up into the vagina with your index finger. **Check to be sure the inner ring is up just past your pubic bone.** (Fig. E)
- You will feel your pubic bone by curving your index finger when it is a couple of inches inside the vagina.
- This step may be hard to do on the first or second try because the FC2 Female Condom is lubricated.
- **Take your time** and push the closed end of the FC2 Female Condom up to where you can feel the pubic bone.
- Make sure the FC2 Female Condom is inserted straight (not twisted) into the vagina.




Fig. E

- It is also important that the outside ring lies against the outer lips of the vagina as shown in Fig. F.
- About one inch of the open end will stay outside your body. Fig. F. While this may look unusual, this part of the FC2 Female Condom helps protect you and your partner during sex. Once the penis enters, the vagina will expand and the slack will decrease.


Fig. F
Outside Ring Correctly Covering Up Area for Sex

5 NEXT STEPS


Fig. G

- Until you and your partner become comfortable using the FC2 Female Condom, use your hand to guide the penis into the vagina. See Fig. G.
- After two or three times, you should become familiar with using the FC2 Female Condom and should hardly notice the pouch or the outer ring during sex.

6 DURING INTERCOURSE

- You may notice that the FC2 Female Condom moves around during sex. See Fig. H.




Fig. H *Fig. I* *Fig. J*

- Movement side-to-side of the outer ring is normal.
- Sometimes the FC2 Female Condom may slip up and down in the vagina, "riding" on the penis. Both the vagina and the penis are still covered.
- However, if you notice the FC2 Female Condom is slipping out of the vagina, **STOP**, remove, and insert a new FC2 Female Condom.
- If you begin to feel the outer ring being pushed into the vagina, **STOP** (Fig. I). Or, if the penis starts to enter underneath or beside the pouch, **STOP** (Fig. J). Take out the FC2 Female Condom. Put in a new FC2 Female Condom. Make sure the outside part lies over the vagina.

7 AFTER INTERCOURSE

- To take out the FC2 Female Condom, squeeze and twist the outer ring to keep the sperm inside the pouch. Pull out gently. Throw away in a trash can. Do not flush. Do not reuse. Please see Fig. K.




Fig. K

REMEMBER

To help reduce your risk of pregnancy and the spreading of a sexually transmitted infection:

- Use a new FC2 Female Condom correctly every time you have sex.
- Follow the directions carefully.
- Be sure you don't tear the FC2 with fingernails or other sharp objects.

QUESTIONS YOU MAY HAVE ABOUT THE CORRECT USE OF THE FC2 FEMALE CONDOM

Q. Will the FC2 Female Condom always protect against pregnancy and STIs?

A. No method is 100 percent effective. Natural rubber latex condoms for men are highly effective at preventing sexually transmitted infections, including AIDS (HIV infection), if used correctly. If you are not going to use a male condom, you can use the FC2 to help protect yourself and your partner.

Q. How do I know when the FC2 Female Condom is up far enough?

A. Using your index finger, push the device so that the lower ring of the device is against the pubic bone. You will find the pubic bone by curving your index finger when it is a couple of inches inside the vagina.

Q. What do I do if the outer ring is pushed inside the vagina?

A. **STOP.** Remove the FC2 Female Condom device. Insert a new FC2 Female Condom according to the directions.

Q. Will I feel the FC2 Female Condom once it is in place?

A. Yes, it is possible that you will be aware of it.

Q. What do I do if the penis is inserted outside the female condom?

A. You should ask your partner to withdraw the penis, and help him reinsert the penis inside the condom.

Q. Will the FC2 Female Condom rip or tear while I am using it?

A. Studies show that the FC2 Female Condom rips or tears less than 1% of the time. If you think the FC2 Female Condom has been ripped or torn, remove it right away, throw it away, and insert a new FC2 Female Condom.

Q. Will the FC2 Female Condom bunch up inside the vagina?

A. The FC2 Female Condom should not bunch up inside if it is inserted correctly. If you feel the outer ring begin to slip inside, **STOP**. Remove the FC2 Female Condom device. Insert a new FC2 Female Condom device according to the directions.

Q. What does the outer ring feel like during sex?

A. While aware of the outer ring, most women say that it does not interfere with intercourse.

Q. What do I do if the FC2 Female Condom does not stay in place during sex?

A. If the FC2 Female Condom moves down the vagina causing discomfort, either push it back up or remove the FC2 Female Condom. If you remove the FC2 Female Condom, use a new one.

DO YOU HAVE A QUESTION ON HOW TO USE THE FC2 FEMALE CONDOM?
Call your healthcare provider, or call 1-800-274-6601.
You can also email us at info@fc2.us.com.

Este folleto explica cómo utilizar el anticonceptivo fc2 FEMALE CONDOM® (CONDÓN FEMENINO FC2) y responde a algunas preguntas que las mujeres a menudo se hacen respecto a su uso.

Lea el presente folleto atentamente antes de usar el CONDÓN FEMENINO fc2 por primera vez.

El Condón Femenino fc2 está indicado para prevenir el embarazo, el VIH/SIDA y otras infecciones de transmisión sexual (ITS). El Condón Femenino FC2 es un condón interno de un solo uso.

INFORMACIÓN IMPORTANTE

Los condones de látex de caucho natural para hombres son altamente efectivos en la prevención de infecciones de transmisión sexual, incluido el VIH/SIDA, de ser usados correctamente.

Si usted no va a utilizar un condón de látex de caucho natural masculino, puede usar el Condón Femenino FC2 para proteger a usted y a su pareja.

El Condón Femenino FC2 sólo funciona cuando se lo utiliza. Utilícelo de manera correcta cada vez que tenga relaciones sexuales.

Antes de probar el Condón Femenino FC2, asegúrese de leer las instrucciones y aprenda a utilizarlo correctamente.

Fabricado por:
Veru Inc.

2916 N. Miami Avenue, Suite 1000
Miami, FL 33127
www.fc2.us.com

FC2 Female Condom es una marca registrada de Veru Inc.

Este producto está protegido por una U.S. 7,726,316 B1, Pat. 8,100,128 y otros.

G0073 Fecha: noviembre de 2021

LEA LA SIGUIENTE INFORMACIÓN ANTES DE UTILIZAR EL CONDÓN FEMENINO fc2

El Condón Femenino fc2 está indicado para prevenir el embarazo, el VIH/SIDA y otras infecciones de transmisión sexual (ITS). El Condón Femenino FC2 es un condón interno de un solo uso. Si se encuentra en tratamiento por una infección de transmisión sexual (ITS) o cree que puede padecer de una, consulte a su proveedor de salud.

PUNTOS IMPORTANTES PARA TENER EN CUENTA

Puntos importantes para tener en cuenta para la mejor protección con el Condón Femenino FC2

1. Antes de probar el FC2, asegúrese de leer las instrucciones y aprenda a utilizarlo correctamente.
2. Practique colocar el FC2 dentro de la vagina sin tener sexo para familiarizarse con el modo de uso.
3. Asegúrese de que el FC2 esté en su lugar antes del ingreso del pene.
4. Ninguno de los estudios realizados demostraron si los espermicidas son compatibles con el material del FC2 o si el uso de un espermicida aumenta la eficacia del FC2.
5. Conserve el FC2 a temperatura ambiente.
6. A continuación se incluyen algunos problemas que se han registrado al utilizar el FC2 durante la relación sexual:
   - El pene puede empujar el condón FC2 (incluido el anillo exterior) adentro de la vagina;
   - El pene se desliza hacia el lateral del FC2 durante la relación sexual e ingresa adentro de la vagina sin la protección del FC2;
   - El Condón Femenino FC2 se rompe durante el uso; y
   - El FC2 se desliza completamente fuera de la vagina.

   En cada uno de los casos anteriores, el riesgo de exposición de la vagina al semen se incrementa. Si alguno de ellos sucede, deberá desechar ese condón y utilizar uno nuevo.
7. No se han realizado estudios para determinar la eficacia del FC2 en la prevención del VIH/SIDA y otras infecciones de transmisión sexual (ITS) cuando se utiliza por vía anal.

UTILICE FC2

- El Condón Femenino FC2 se inserta antes de la relación sexual. La mayoría de las mujeres insertan el FC2 entre 2 a 20 minutos antes de esta.
- El Condón Femenino FC2 debe retirarse luego del acto sexual y antes de ponerse de pie. Sólo puede utilizarse una vez.
- El FC2 puede utilizarse con la mayoría de los lubricantes sexuales de base mineral, de silicona y de agua, disponibles en el mercado.

BENEFICIOS DE LA UTILIZACIÓN DEL FC2

- El uso del FC2 durante la relación sexual puede colaborar en la protección contra las infecciones de transmisión sexual o a evitar el embarazo.
- El FC2 no contiene látex de caucho natural.
- El FC2 ofrece a las mujeres una opción para protegerse que ellas pueden iniciar.
- Las mujeres pueden insertar el FC2 antes del acto sexual. Esto permite que el acto sea más espontáneo, dado que no es necesaria la erección para insertar el FC2.

LOS PELIGROS DE UTILIZAR EL FC2


- Utilice el Condón Femenino FC2 correctamente cada vez que tenga relaciones sexuales. Si el FC2 no se usa del modo correcto, el riesgo de contraer infecciones de transmisión sexual o de quedar embarazada se incrementará.
- El uso incorrecto podría ocasionar que el FC2 no ofrezca protección alguna. Para evitar el riesgo de uso incorrecto, lea las instrucciones atentamente antes de usarlo.
- Las investigaciones demuestran que los siguientes hechos son raros pero pueden ocurrir al utilizar el FC2:
  - incomodidad durante la inserción
  - dolor luego de la inserción, antes de la relación sexual
  - incomodidad durante la relación sexual
  - sensación de ardor, salpullido o comezón.

ALGUNAS CARACTERÍSTICAS DE USAR EL CONDÓN DE LA HEMBRA FC2

- Usted misma puede colocarse el Condón Femenino FC2. Ésto le ofrece la oportunidad de protegerse por sí sola.
- El Condón Femenino FC2 se calentará apenas lo inserte. Es fuerte y suave.

DESCRIPCIÓN

El Condón Femenino FC2 es una cubierta suave y que recubre en forma suelta y tiene un anillo en cada extremo. El anillo interno del extremo cerrado está compuesto de poliuretano y se utiliza para ayudar a insertar el Condón Femenino FC2 dentro de la vagina y ayuda a mantenerlo en su lugar. No está conectado con el condón femenino. El borde exterior de la cubierta del extremo abierto se mantiene fuera de la vagina. El Condón Femenino FC2 viene prelubricado con silicona.



SOBRE LA EFECTIVIDAD DEL FC2

La efectividad del Condón Femenino FC2 se basa en un estudio realizado en Sudáfrica en el que se compar el Condón Femenino FC2 con una versión anterior, el Condón Femenino FC1[1]. Este estudio comparativo analizó el FC2 y el FC1 respecto de otros cuatro posibles problemas descriptos anteriormente en el rótulo:

- el condón femenino puede romperse durante el uso;
- el condón femenino podría deslizarse fuera de la vagina durante el uso;
- el pene podría empujar el condón femenino adentro de la vagina; y
- el pene podría empujar el condón hacia adentro de la vagina (sin protección).

Estos cuatro problemas son las únicas formas en las que la vagina podría estar expuesta al semen al utilizar el Condón Femenino FC2 y este estudio comparativo demuestra que las cifras de estos problemas en los condones femeninos FC1 y FC2 son similares. Aunque el FC2 no fue estudiado en cuanto a su anticoncepción ni a la protección frente a las ETS, la efectividad del FC2 puede inferirse sobre la base de este estudio comparativo y un estudio anterior que demuestra la efectividad del FC1 para anticoncepción del FC1 para anticoncepción. Éstos resultados se consignan en la siguiente tabla.

A continuación se incluye información sobre el modo en el que diferentes métodos anticonceptivos la protegen frente al embarazo. Las cifras se basan en el uso habitual y perfecto. Las cifras de uso habitual indican que método podría haberse utilizado de manera incorrecta, podría no haberse utilizado en todo acto sexual o podría haber fallado durante el uso. Las cifras de uso perfecto indican que el método fue utilizado correctamente con cada acto sexual, pero falló, de todos modos.

% DE MUJERES QUE EXPERIMENTARON UN EMBARAZO NO BUSCADO DENTRO DEL PRIMER AÑO DE USO

MÉTODO ANTICONCEPTIVO	USO HABITUAL	USO PERFECTO
Ningún método	85	85
Espermicidas	29	18
Retiro	27	4
Métodos basados en la concienciación sobre Fertilidad	25	–
Método de barrera		
Capuchón cervical		
Sin partos previos	16	9
Partos previos	32	20
Diafragma + espermicida	16	6
Condón-		
Condón femenino FC1	21	5
Condón masculino	15	2
Métodos hormonales		
Píldoras combinadas y píldoras sólo a base de progestina	8	0.3
Evra Patch®	8	0.3
NuvaRing®	8	0.3
Depo-Provera®	3	0.3
Implanon®	0.05	0.05
IUD		
ParaGard® (T de cobre)	0.8	0.6
Mirena® (DIU-LNG)	0.2	0.2
Esterilización		
Esterilización femenina	0.5	0.5
Esterilización masculina	0.15	0.10

Píldoras anticonceptivas de emergencia: El tratamiento iniciado dentro de las 72 horas posteriores a la relación sexual sin protección disminuye el riesgo de embarazo por lo menos en un 75%.

Método de amenorrea de la lactancia: el MELA es un método anticonceptivo altamente efectivo y temporal.

Fuente: Trussell J. Contraceptive efficacy. In Hatcher RA, Trussell J, Nelson AL, Cates W, Stewart FH, Kowal D. Contraceptive Technology: Nineteenth Revised Edition. New York NY: Ardent Media, 2007; www.contraceptivetechnology.org

1 Referencia: M. Beksinska et al. Performance of the Reality polyurethane female condom and a synthetic latex prototype: a random crossover trial among South African women. Contraception. 2006 Apr; 73(4):386-93.

CÓMO UTILIZAR EL CONDÓN FEMENINO fc2

1 PARA ABRIR EL PAQUETE



- Encuentre la flecha en la lengüeta de apertura en la esquina derecha superior y tire para abrir (Fig. A).
- Frote la parte externa del condón femenino para asegurarse de que el lubricante está distribuido uniformemente dentro de la bolsa desde la parte inferior a la superior.
- Saque el FC2 y mírelo de cerca para asegurarse de que no esté dañado.


2 PARA INTRODUCIR EL CONDÓN FEMENINO FC2




- Encuentre una posición cómoda. Intente pararse con un pie sobre una silla, siéntese con las rodillas separadas o ubíquese en cuclillas.
- Asegúrese de que el anillo interno está en el extremo cerrado inferior de la bolsa (Fig. B).

3 PRÓXIMOS PASOS

- Sujete el condón femenino con el extremo abierto hacia abajo. Al sujetar la parte externa de la bolsa, apriete el anillo interno con sus dedos pulgar y medio.
- Coloque su dedo índice entre el pulgar y el dedo medio y siga apretando el anillo interno (Fig. C (a)). O sólo apriete, lo que le sea más fácil (Fig. C (b)).
- Mientras está apretando el Condón Femenino FC2, con la otra mano separe los labios de la vagina como si se estuviera preparando para introducir un tampón.
- Introduzca el Condón Femenino FC2 como se muestra en la Fig. D.







Tómese su tiempo. Si el Condón Femenino FC2 está resbaloso para introducirlo, suéltelo y comience de nuevo.

4 PRÓXIMOS PASO

- Ahora empuje el anillo interno y la bolsa el tramo que queda dentro de la vagina con su dedo índice. **Verifique que el anillo interno esté por encima de su pubis.** (Fig. E)



- Encontrará su pubis haciendo una curva con su dedo índice cuando esté unas pulgadas dentro de la vagina.
- Este paso puede ser difícil la primera o la segunda vez porque el Condón Femenino FC2 está lubricado.
- **Tómese su tiempo** y empuje el extremo cerrado del Condón Femenino FC2 hasta donde pueda sentir el pubis.
- Asegúrese de insertar el Condón Femenino FC2 derecho (y no torcido) adentro de la vagina.
- Es importante también que el anillo externo esté fuera de los labios externos de la vagina como se muestra en la Fig. F.
- Alrededor de una pulgada del extremo abierto quedará fuera de su cuerpo. Fig. F. Si bien esto puede parecer poco usual, es esta parte del Condón Femenino FC2 la que ayuda a protegerla a usted y a su pareja durante el sexo. Una vez que el pene se introduce, la vagina se expande y no quedará ya suelto el condón.




5 PRÓXIMOS PASOS



- Hasta que usted y su pareja se familiaricen con el Condón Femenino FC2, use su mano para guiar al pene hacia la vagina. Vea la Fig. G.
- Luego de usarlo dos o tres veces, ya se habrán familiarizado con el Condón Femenino FC2 y apenas notarán la bolsa o el anillo externo durante la relación sexual.


6 DURANTE LA RELACIÓN SEXUAL

- Quizás note que el Condón Femenino FC2 se mueve durante la relación sexual. Vea la Fig. H.



- El movimiento hacia los lados del anillo externo es normal.
- En ocasiones puede que el Condón Femenino FC2 se deslice hacia afuera y adentro de la vagina, junto con el pene. Tanto la vagina como el pene siguen cubiertos.
- Sin embargo, si nota que el Condón Femenino FC2 se está saliendo de la vagina, **DETÉNGASE**, quítelo y colóquese uno nuevo.
- Si comienza a sentir que el anillo externo se está introduciendo en la vagina (Fig. I), o si el pene comienza a entrar por debajo o por el costado de la bolsa, **DETÉNGASE** (Fig. J). Sáquese el Condón Femenino FC2 y colóquese uno nuevo. Asegúrese de que la parte externa quede por fuera de la vagina.

7 LUEGO DE LA RELACIÓN SEXUAL



- Para sacarse el Condón Femenino FC2, apriete y retuerza el anillo externo para mantener el esperma para mantener el esperma dentro de la bolsa. Sáquelo suavemente. Arrójelo a un cesto de basura. No lo tire por el inodoro. No lo vuelva a usar. Vea la Fig. K.

RECUERDE

Para ayudar a reducir el riesgo de embarazo y el contagio de infecciones de transmisión sexual:

- Utilice el Condón Femenino FC2 correctamente cada vez que tenga relaciones sexuales.
- Siga las instrucciones atentamente.
- Asegúrese de no romper el condón FC2 con sus uñas ni con otros objetos cortantes.

POSIBLES PREGUNTAS SOBRE EL USO CORRECTO DEL CONDÓN FEMENINO fc2

P. ¿El Condón Femenino FC2 siempre ofrece protección contra el embarazo y las ITS?

R. Ningún método es cien por ciento efectivo. Los condones de látex de caucho natural para hombres son altamente efectivo en la prevención de enfermedades de transmisión sexual, incluido el SIDA (infección de VIH), de ser usado correctamente. Si usted no va a utilizar un condón de látex de caucho natural masculino, puede usar el Condón Femenino FC2 para proteger a usted y a su pareja.

P. ¿Cómo sé cuándo el Condón Femenino FC2 está ubicado en la profundidad correcta?

R. Con el dedo índice empuje el dispositivo de manera que el anillo inferior se ubique contra el pubis. Encontrará el pubis curvando su dedo índice cuando esté a pocas pulgadas dentro de la vagina.

P. ¿Qué hago si el anillo externo ingresa dentro de la vagina?

R. DETÉNGASE. Retire el Condón Femenino FC2 e inserte uno nuevo según las instrucciones.

P. ¿Sentiré el Condón Femenino FC2 una vez que esté ubicado en su lugar?

R. Sí, es posible que lo note.

P. ¿Qué hago si el pene se inserta fuera del condón femenino?

R. Debe pedirle a su pareja que lo retire y la ayude a reinsertarlo dentro del condón.

P. ¿El Condón Femenino FC2 se rasgará o se romperá mientras lo estoy usando?

R. Los estudios demuestran que el Condón Femenino FC2 se rasga o rompe en menos de un 1% de las veces. Si cree que el Condón Femenino FC2 se ha rasgado o roto, retírelo de inmediato, deséchelo e inserte uno nuevo.

P. ¿El Condón Femenino FC2 se amontonará dentro de la vagina?

R. El Condón Femenino FC2 no debería amontonarse dentro si se inserta de manera correcta. Si siente que el anillo exterior comienza a deslizarse hacia adentro, **DETÉNGASE.** Retire el Condón Femenino FC2 e inserte uno nuevo según las instrucciones.

P. ¿Cómo se siente el anillo exterior durante la relación sexual?

R. Aunque sienten el anillo exterior, la mayoría de las mujeres afirman que no interfiere con la relación sexual.

P. ¿Qué hago si el Condón Femenino FC2 no se mantiene en su lugar durante la relación sexual?

R. Si el Condón Femenino FC2 se mueve hacia afuera de la vagina y causa incomodidad, empújelo hacia adentro o retire el Condón Femenino FC2. Si lo hace, utilice uno nuevo.

¿TIENE ALGUNA PREGUNTA SOBRE EL MODO DE USO DEL CONDÓN FEMENINO FC2?
Llame a su proveedor de salud o comuníquese al 1-800-274-6601.

También puede enviarnos un correo electrónico a info@fc2.us.com.